



Certificate of Compliance

Certificate: 1088398

Master Contract: 204767

Project: 1800502

Date Issued: 2006/08/23

Issued to: Camille Bauer AG

Aargauerstrasse 7
Wohlen, 5610
Switzerland
Attention: Mr. Markus Fahndrich

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Eknath Desai, C.E.T.

Authorized by: Renzo Pupulin, C.E.T., Product Group Manager

PRODUCTS

CLASS 8721 84 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

CLASS 8721 04 - LABORATORY EQUIPMENT - Electrical

Programmable Multi-Transducer, Model DME 4XX* Series, and DMT10XX** Series, Power Supply: AC/DC 85-230V or AC/DC 24-60V, 50/60Hz, 9W, Permanently Connected, Continuous Operation, Class II, Pollution Degree 2, Installation Category II.

The *'s in the model designation may be any number 0 to 9 or letter A to Z or blank denoting SELV/ELVEL secondary circuits or minor mechanical differences.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1088398

Master Contract: 204767

Project: 1800502

Date Issued: 2006/08/23

APPLICABLE REQUIREMENTS

CAN / CSA - C22.2 No 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std No 3101.1 - Electrical Equipment for Laboratory Use; Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 1088398

Master Contract: 204767

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
1800502	2006/08/23	UPDATE REPORT # 1188398 TO COVER MINOR ALTERNATE CONSTRUCTION.

History

1088398 - January 25, 2001 - Original certification.

Supplement Notes

1313258 - April 26, 2002 - Alternate construction.